

CHEMISTRY PUMPING UNIT SERIES

PC 3001 VARIO select
PC 3001 VARIO select TE
PC 3001 VARIO select IK
PC 3001 VARIO select EKP



Instructions for use



Original instructions

Keep for future use!

This document may only be used and distributed in its complete and original form. It is the user's responsibility to ensure the validity of this document with respect to the product.

Manufacturer:

VACUUBRAND GMBH + CO KG

Alfred-Zippe-Str. 4

97877 Wertheim

GERMANY

Head office: +49 9342 808-0

Sales: +49 9342 808-5550

Service: +49 9342 808-5660

Fax: +49 9342 808-5555

E-mail: info@vacuubrand.com

Web: www.vacuubrand.com

*Thank you for purchasing this product from **VACUUBRAND GMBH + CO KG**. You have chosen a state-of-the-art, high-quality product.*

TABLE OF CONTENTS

1	About this manual	6
1.1	User information	6
1.2	Instruction manual layout	7
1.3	Presentation conventions.....	8
1.4	Symbols and pictograms	9
1.5	Action instructions	10
1.6	Abbreviations	10
1.7	Explanation of terms.....	12
2	Safety instructions	13
2.1	Use	13
2.1.1	Proper use.....	13
2.1.2	Improper use	14
2.1.3	Foreseeable misuse.....	14
2.2	Responsibilities	15
2.3	Target group description	15
2.4	General safety information.....	16
2.5	Protective clothing.....	16
2.6	Safety measures	17
2.7	Laboratory and work materials.....	17
2.8	Possible sources of danger	19
2.9	Motor protection	21
2.10	ATEX device category	22
2.11	Disposal	23
3	Product description	24
3.1	Basic configuration of pumping unit series.....	24
3.2	Chemistry pumping unit series	25
3.3	Condensers and coolers	26
3.3.1	Separator/condenser at inlet	26
3.3.2	Condenser at outlet.....	27
3.4	Application example	29
4	Installation and connection	31
4.1	Transport	31
4.2	Installation.....	32
4.3	Connection (supply connections)	33
4.3.1	Vacuum connection (IN).....	33

4.3.2	Exhaust connection (OUT)	35
4.3.3	Coolant connection on the condenser	36
4.3.4	Dry ice condenser	37
4.3.5	Venting connection	39
4.3.6	Gas ballast (GB)	42
4.4	Electrical connection	43
5	Operation	46
5.1	Switching on	46
5.2	Operation with controller	47
5.2.1	Operator interface	47
5.2.2	Operation	49
5.2.3	Operation with gas ballast	50
5.3	Switching off (decommissioning)	51
5.4	Storage	52
6	Troubleshooting	53
6.1	Technical assistance	53
6.2	Error – cause – corrective measure	53
7	Cleaning and maintenance	56
7.1	Information on service activities	57
7.2	Cleaning	59
7.2.1	Housing surface	59
7.2.2	Emptying the glass flask	60
7.2.3	Clean the sensor and venting valve	60
7.2.4	Cleaning or replacing PTFE molded hoses	64
7.3	Vacuum pump maintenance	64
7.3.1	Maintenance items	64
7.3.2	Change the diaphragms and valves	66
8	Annex	78
8.1	Technical data	78
8.2	Wetted materials	80
8.3	Rating plate	81
8.4	Order data	82
8.5	Service information	83
8.6	EU declaration of conformity	85
8.7	UKCA conformity declaration	86
8.8	Certificate (CUS)	87

1 About this manual

This User manual is part of the pumping unit you have purchased. This instruction manual is valid for all pumping unit models, together with the instruction manual of the **VACUU SELECT** controller, and is intended specifically for operators.

1.1 User information

Safety

Instruction manual
and safety

- Read the User manual carefully before using the product.
- Store the User manual in a place where it is accessible and close at hand at all times.
- Correct use of the product is essential for safe operation. Above all else, please follow the safety instructions!
- In addition to the information in this User manual, please also observe the applicable national regulations on accident prevention and occupational safety.

General

General information

- When passing this product on to third parties, please also include the User manual.
- All figures and drawings are examples and are solely intended for the purpose of better understanding.
- We reserve the right to make technical changes in the course of continuous product improvement.
- In the interest of readability, the Pumping unit is equally used in place of the product name Chemistry pumping unit PC 3001 VARIO select.

Copyright

Copyright® and
copyright law

The content of this User manual is protected by copyright law. Copies for internal purposes are permitted, e.g. for trainings. © **VACUUBRAND GMBH + CO KG**

Contact

Please contact us

- In case of an incomplete User manual, you can request a replacement from us. Alternatively, our download portal is available to you: www.vacuubrand.com
- Call us or write to us if you have any other questions about the product, need additional information or want to give us feedback on the product.
- When you contact our service department, please have the serial number and product type on hand --> see the nameplate on the product.


1.2 Instruction manual layout

Manual organization

The instruction manual for the pumping unit, the controller and available accessories is organized in a modular format; this means that the instructions are divided into individual, separate instruction brochures.

Instruction modules

Pumping unit series and modular instruction manuals



Meaning

- 1** Safety instructions for vacuums
- 2** Instruction manual: Vacuum controller – Control and operation
- 3** Instruction manual: Pumping unit – Connection, operation, maintenance, mechanical system


4 Optional instruction manual: Accessories


1.3 Presentation conventions

Warning messages

Presentation of warning messages

	DANGER
	<p>Warning of imminent danger.</p> <p>Failure to observe this warning may result in imminent danger to life or severe injury.</p> <ul style="list-style-type: none"> ➤ Please follow the instructions for prevention!


	WARNING
	<p>Warning of a potentially dangerous situation.</p> <p>Failure to observe this warning may result in danger to life or serious injury.</p> <ul style="list-style-type: none"> ➤ Please follow the instructions for prevention!

	CAUTION
	<p>Indicates a potentially dangerous situation.</p> <p>Failure to observe this caution may result in minor injuries or material damage.</p> <ul style="list-style-type: none"> ➤ Please follow the instructions for prevention!

NOTICE	
<p>Reference to a potentially harmful situation.</p> <p>Failure to observe this note may result in material damage.</p>	

Additional information

Presentation of information and tips

	<p>General information about:</p> <ul style="list-style-type: none"> ⇒ Tips and tricks ⇒ Helpful functions or activities
---	---

1.4 Symbols and pictograms

This instruction manual uses symbols and pictograms. These safety symbols and pictograms indicate specific dangers or requirements when handling the product. Warning signs with safety symbols on the product provide a visualization of the potential hazard.









Safety symbols

Explanation of safety symbols

	General danger sign.		Warning of electrical voltage.
	Warning of hot surface.		Electrostatically sensitive components ESD.
	General mandatory action sign.		Disconnect power plug from electrical outlet.

Other symbols and pictograms

Additional symbols

	Positive example – Right! Result – o.k.		Negative example – Wrong!
	Reference to content in this User manual.		Reference to content in supplementary documents.
	Ensure sufficient air circulation.		
	Electrical and electronic equipment and batteries must not be disposed of in household waste at the end of their service life.		
	Inlet current arrow – vacuum connection		
	Outlet current arrow – exhaust gas		

1.5 Action instructions

Action instructions (simple)

- Action instructions
- ⇒ You are requested to take action.
 - Result of the action



Action instructions (multiple steps)

1. First action step
 2. Next action step
- Result of the action

Action instructions that require several steps must be followed in the order they are described.

Action instructions (image description)


-> Example
Principle presentation
Operating steps
presented in images

<div data-bbox="438 963 502 1019" style="border: 1px solid black; padding: 2px; width: 30px; margin: 0 auto;">1</div> 	<div data-bbox="981 963 1045 1019" style="border: 1px solid black; padding: 2px; width: 30px; margin: 0 auto;">2</div> 
<p>1. First action step.</p>	<p>2. Next action step.</p> <ul style="list-style-type: none"> <input checked="" type="checkbox"/> Interim result or result of the action

1.6 Abbreviations

Abbreviations used

>/	No greater than
abs.	Absolute
AK	Separator flask
ATM	Atmospheric pressure (bar graph, program)
di	Diameter
DN	Diameter nominal
EK	Emission condenser

EKP	Peltronic emission condenser or Peltronic EK
EX ¹	Outlet (exhaust, exit), exhaust gas connection
	ATEX device marking
FPM	Fluoropolymer rubber
Gas type ind.	Gas type independent
GB	Gas ballast
If necessary	If necessary
Size	Size
IK	Immission condenser
IN ¹	Inlet, vacuum connection
KF	Small flange
Max.	Maximum value
Min.	Minimum value
w/o EK	Without emission condenser
PA	Polyamide
PBT	Polybutylene terephthalate
PC ...	Chemistry pumping unit with type code
PE	Polyethylene
RMA No.	Return authorization number
so-called	so-called
SW	Wrench size (tool)
TE	Dry ice condenser
resp.	Responsible person(s)
e.g.	For example

¹ Labeling on vacuum pump or component, see also product specific abbreviations under:
→ **Chemistry pumping unit series on page 25**

1.7 Explanation of terms

Product-specific terms

Separator flask	Glass flask/separator installed at the inlet or outlet.
Emission condenser²	Cooling condenser with receiving flask installed at the outlet (pressure side).
Fine vacuum	Pressure measurement range in the vacuum technology, from: 1 mbar–0.001 mbar (0.75 Torr–0.00075 Torr)
Rough vacuum	Pressure measurement range in the vacuum technology, from: Atmospheric pressure–1 mbar (atmospheric pressure–0.75 Torr)
Immission condenser²	Cooling condenser with receiving flask installed at the inlet (vacuum side).
PC 3001 VARIO select	Vacuum pumping unit with speed control for precise vacuum regulation with VACUU SELECT controller and VACUU·SELECT sensor.
Peltronic	Electronic cooler with Peltier elements installed at the outlet (pressure side); condenses solvent vapors without external cooling medium.
Dry ice condenser²	Cooling condenser with receiving flask and dry ice as cooling medium installed at the outlet (pressure side).
VACUU·BUS	VACUUBRAND bus system for the communication of peripheral devices with VACUU BUS-capable measuring equipment and controllers.
VACUU·BUS address	Address that enables a clear assignment of the VACUU BUS client in the bus system, e.g. for the connection of several sensors of the same measurement range.
VACUU BUS client	Peripheral device or components with VACUU BUS connection, which is integrated in the bus system, e.g. sensors, valves, level indicators, etc.
VACUU BUS plug	4-pin round plug for the VACUUBRAND bus system.
VACUU BUS configuration	Using a measuring device or controller to assign a new VACUU BUS address to a VACUU BUS component.
VACUU·SELECT	Vacuum controller, controller with touchscreen; consisting of control unit and vacuum sensor.
VACUU·SELECT sensor	Vacuum sensor with integrated venting valve.
VARIO drive	Speed control for vacuum pumps; the motor only runs as fast as needed.

² Only suitable for the condensation of vapors.

2 Safety instructions

The information in this chapter must be observed by all persons who work with the device described here.

The safety instructions are valid for all life stages of the product.

2.1 Use

The device may only be used in perfect technical condition.

2.1.1 Proper use

Proper use A chemistry pumping unit of the PC 3001 VARIO select product series is a vacuum system consisting of a vacuum pump, controller, vacuum sensor and separator for generating and regulating a rough vacuum in installations designed for this purpose.

Attached coolers (emission condenser, immission condenser, dry ice cooler, Peltronic emission condenser), including separator and flask, are exclusively designed for condensing vapors.

Application example: Evacuating distillation instruments, especially rotary evaporators.

The vacuum system should only be used in a dry, non-explosive environment.

Proper use also includes:

- following the instructions in the document *Safety instructions for vacuums*,
- observing the instruction manual,
- observing the instruction manual of connected components,
- complying with inspection and maintenance intervals and having this carried out only by qualified personnel.
- using only approved accessories or replacement parts.

Any other use or use beyond this is considered improper.

2.1.2 Improper use

Improper use Improper use or any use that does not correspond with the technical data can lead to personal or material damage.

Improper use is considered:

- use that contradicts the proper use,
- operation in unauthorized ambient and operating conditions,
- operation with obvious faults, damages or defective safety equipment,
- unauthorized extensions and modifications, especially when these compromise safety,
- use in an incomplete state,
- operation with sharp-edged objects,
- pulling connectors on the cable out of the socket,
- vacuuming, pumping and condensing solid materials or liquids.

2.1.3 Foreseeable misuse

Misuse In addition to improper use, there are other types of use that are forbidden when handling the device.

Forbidden types of use are primarily:

- use on people or animals,
- setup and operation in a potentially explosive environment,
- use in mining or underground,
- using the product to generate pressure,
- fully exposing vacuum equipment to the vacuum,
- submerging vacuum equipment in liquids, exposing to spray water or steam spraying,
- pumping oxidizing and pyrophoric materials, liquids or solids,
- pumping media that is hot, unstable, potentially explosive or explosive,
- pumping materials that can react explosively under impact and/or increased temperature without air supply.

The user must prevent the penetration of foreign bodies, hot gases and flames.

2.2 Responsibilities

Follow the instructions for all actions as they are specified in this instruction manual.

Responsibilities of the operator

Operator responsibilities

The operator defines the responsibilities and ensures that only trained or qualified personnel work on the vacuum system. This applies in particular to connection, assembly work, maintenance tasks and fault elimination.

Users in the competency areas listed in the → **Target group description on page 15** must have the corresponding qualification for the listed activities. Only qualified electricians are permitted to carry out special work on electrical equipment.

Personnel responsibilities

Personnel responsibilities

For activities that require protective clothing, the personal protection equipment specified by the operator must be worn.

When the vacuum system is not in proper operating condition, it must be secured against accidental restart.

- ⇒ Always work with safety in mind.
- ⇒ Follow the operator’s instructions and the national regulations on accident prevention and occupational safety.



Personal conduct can contribute to the prevention of occupational accidents.

2.3 Target group description

Target groups

The instruction manual must be read and observed by every person entrusted with one of the following activities.

Personnel qualification

Qualification description

Operator	Laboratory personnel, e.g. chemists, physicists, lab technicians
Qualified employee	Person with professional qualification for maintenance and/or repair in the field of: mechanical systems, electrical systems or laboratory equipment. The assigned work can be assessed and potential dangers detected.

Responsible specialist Qualified employee with additional field, department or division responsibility who is assigned by the operator.

Responsibility matrix

Who-does-what matrix

Task	Operator	Qualified employee	Responsible specialist
Installation	x	x	x
Initial operation	x	x	x
Network integration			x
Operation	x	x	x
Fault reporting	x	x	x
Fault elimination	(x)	x	x
Device fuse replacement		x	x
Maintenance		x	x
Repair ³		x	x
Repair order			x
Cleaning, simple	x	x	x
Emptying separator	x	x	x
Decommissioning	x	x	x
Decontamination ⁴		x	x

2.4 General safety information

Quality standards and safety


Products of **VACUUBRAND GMBH + CO KG** are subject to high quality control requirements in terms of safety and operation. Each product is put through an extensive test program before delivery.

Follow the instructions for all actions as they are specified in this instruction manual.

2.5 Protective clothing

Special protective clothing is not required for operation of the vacuum pump. Follow the operator’s instructions for your workplace.

For cleaning, maintenance and repair work, we recommend wearing fully adequate protective gloves, clothing and glasses.



³ See also homepage: VACUUBRAND > Support > [Repair instructions](#)

⁴ Or have decontamination carried out by a qualified service provider.

- ⇒ Wear your personal protection equipment when handling chemicals.

2.6 Safety measures

Manufacturer measures

Products of **VACUUBRAND GMBH + CO KG** are subject to high quality control requirements in terms of safety and operation. Each product is put through an extensive test program before delivery.

Operator measures

Operator measures

- ⇒ Use your vacuum only when you have understood the instruction manual and the operation.
- ⇒ Replace defective components immediately, e.g. broken power cables, defective hoses or flasks.
- ⇒ Use only original accessories and components that are designed for the vacuum technology, e.g. vacuum hose, separator, vacuum valve, etc.
- ⇒ When handling contaminated parts, follow the relevant regulations and protection measures; this also applies to sending parts in for repair.
- ⇒ For repairs, send us the carefully filled out and signed **Clearance Certificate** **before** you send your product in for repair. Hazardous materials must be able to be excluded for all repair shipments to our service department.

2.7 Laboratory and work materials



DANGER

Hazardous materials leak at the outlet.

When operating the vacuum, hazardous, toxic substances can leak into the ambient air at the outlet.

- Please observe the safety regulations for handling hazardous materials and media.
- Remember that adhesive process media can present a danger to humans and the environment.
- Install and use suitable separators, filters or extraction equipment.

Dangers posed by different substances

Pumping different substances

Pumping different substances or media can trigger a reaction between materials.

Work materials that leak into the vacuum pump with the gas flow can damage the vacuum pump. Hazardous substances can form deposits in the vacuum pump.

Possible protective measures

Protective measures, depending on the application

- ⇒ Rinse the vacuum pump with inert gas or air before you change the pumping medium.
- ⇒ Use inert gas to dilute critical mixtures.
- ⇒ Prevent the release of hazardous, toxic, explosive, corrosive, health-damaging or environmentally harmful fluids, gases or vapors, e.g. by using suitable laboratory equipment with an extraction system and ventilation control.
- ⇒ Protect the inside of the vacuum pump from deposits or humidity, e.g. by using a gas ballast feed system.
- ⇒ Observe the interactions and possible chemical reactions of the pumped media.
- ⇒ Check the compatibility of the pumped substances with the medium-affected materials of the pumping unit.
- ⇒ Contact us if you have concerns about using your vacuum pump with special work materials or media.

Preventing foreign bodies inside the pump

Observing the vacuum pump design

Particles, liquids and dust are not permitted inside the vacuum pump.

- ⇒ Do not pump any substances that can form deposits inside the vacuum pump.
- ⇒ Install suitable separators and/or filters in front of the inlet. Suitable filters are chemical resistant, clog-free and ensure a constant flow rate.
- ⇒ Replace porous vacuum hoses promptly.

2.8 Possible sources of danger

Considering mechanical stability

Observing mechanical stability

Due to the high compression ratio of the pumps, a higher pressure can build up at the outlet than the mechanical stability of the system allows.

- ⇒ Always ensure that the exhaust line is open and free of pressure. To guarantee an unobstructed emission of gases, the outlet must remain unblocked.
- ⇒ Prevent uncontrolled excess pressure, e.g. due to shut-off or blocked piping system, condensate or clogged exhaust line.
- ⇒ At the gas connections, the connections for inlet (IN) and outlet (EX) must never be confused.
- ⇒ Observe the max. pressure at the inlet and outlet of the pump as well as the max. permissible differential pressure between inlet and outlet, in accordance with the *technical data*.
- ⇒ The system to be evacuated and all hose connections must be mechanically stable.
- ⇒ Attach the coolant hoses to the hose nozzles so that they do not come loose unintentionally.

Preventing condensate return flow

Preventing backlog in the exhaust line

Condensate can damage the pump head. Condensate must never flow back through the hose line into the outlet and into the pump head. Liquid is not allowed to collect in the exhaust line.

- ⇒ Prevent condensate return flow by using a separator. Condensate is not allowed to enter the inner housing through the hose lines.
- ⇒ If possible, lay the exhaust hose so that it is descending from the outlet; i.e. running downward, so that no backlog can form.
- ⇒ Incorrect measurement due to blocked vacuum line, e.g. condensate in the vacuum line can distort the measurements of the vacuum sensor.
- ⇒ Prevent excess pressure in the suction hose ($>/ 1060$ mbar [$>/ 795$ Torr]).

Dangers during ventilation

Looking out for dangers during ventilation

Depending on the process, a potentially explosive mixture can form in installations or other dangerous situations can occur.

Dangers due to residual energy

Possible residual energies

After the vacuum pump has been switched off and disconnected from the power network, dangers may still be present due to residual energies:

- Thermal energy: Motor waste heat, hot surface, compression heat.
- Electrical energy: Installed condensers have a discharge time of up to 3 minutes.

Please note the following before performing actions:

- ⇒ Allow the vacuum pump to cool down.
- ⇒ Wait until the condenser have discharged.

Dangers due to hot surfaces or overheating

Surface temperatures

The surface of vacuum pumps can reach temperatures greater than 70 °C during operation, especially when vacuuming heated media.

- ⇒ Avoid direct contact with the surface or wear heat-resistant safety gloves if contact cannot be excluded.
- ⇒ Provide contact protection if the surface temperature should be elevated on a regular basis.
- ⇒ Allow the vacuum pump to cool down before performing maintenance tasks.

Overheating

The vacuum pump can become damaged due to overheating. Potential causes are insufficient air supply to the cooling fan and/or non-compliance with minimum distances.

- ⇒ When setting up the device, observe a minimum distance of 5 cm between the cooling fan and adjacent parts (e.g. housing, walls, etc.).
- ⇒ Always ensure that there is sufficient air supply; if necessary, provide an external forced ventilation system.
- ⇒ Place the device on a stable surface. A soft surface, e.g. foam as sound absorber, can impair and block the supply of air.
- ⇒ Clean dirty ventilation slots.

- ⇒ Remove the cover hood used as transport protection before you put the device into operation.
- ⇒ Avoid applying too much heat due to hot process gases.
- ⇒ Observe the maximum permissible media temperature according to the *technical data*.

Handling coolants and cryogenic materials properly

Proper handling of cryogenic materials

Cryogenic materials may cause frostbite (cold burns) when contact is made with skin.

- ⇒ Observe the valid regulations for handling cryogenic materials.
- ⇒ Use only permitted transport containers.
- ⇒ Take the necessary safety precautions when handling cryogenic cooling media, e.g. dry ice.
- ⇒ Do not use any damaged components.
- ⇒ Wear your personal protection equipment when handling hazardous materials.
- ⇒ Ensure proper ventilation of the workplace.

Dry ice is not permitted for use in gas-tight containers. Do not fasten the cover on the dry ice condenser. The pressure compensation between the coolant and the atmosphere must be ensured at all times.

Ensure that signs are readable

Labels and signs

Make sure that instructions and signs attached to the device remain in a readable condition:

- ⇒ Labels for connections
- ⇒ Warning and information signs
- ⇒ Motor data and type plates

2.9 Motor protection

Overheating protection, blockage protection

The pump motor has a temperature sensor on the circuit board to protect it against overload. If the temperature is excessive or the motor is blocked, the vacuum pump switches off.

Restart procedure

If the vacuum pump switches off due to these safety measures, the fault must be reset manually: Disconnect the pumping unit from the power supply -> Eliminate the cause of the error -> Restart the pumping unit.

2.10 ATEX device category

Installation and explosive environment

Installation and operation in areas where an explosive atmosphere can occur in dangerous quantities is not permitted.


The user is responsible for assessing the hazards for the device, so that any protection measures can be implemented for the installation and safe operation.

The ATEX certification applies only for the inner area in contact with media of the of the vacuum pump, not for the surrounding area.

ATEX device marking

ATEX device category



Vacuums labeled with the marking  are certified in accordance with the ATEX marking on the type plate.

Operation is only permitted in technically sound condition.

The product is designed for a low degree of mechanical danger and must be installed so that it cannot be mechanically damaged from the outside.

ATEX device category
and peripheral de-
vices

The ATEX device category of the of the vacuum pump depends on the connected components and the periphery. Components and peripheral devices must have the same or higher ATEX classification.

Preventing sources
of ignition


The use of ventilation valves is only permitted if it is ensured that this normally does not produce explosive mixtures in the interior of the of the vacuum pump or that in all probability explosive mixtures are only briefly or rarely produced.

⇒ If necessary, ventilate with inert gas.

Information about the ATEX device category can be found online:
[ATEX information](#)

Restrictions on operating conditions

Explanation of usage
conditions X
Type plate example




Meaning for devices marked with **X**:

- The devices have a low mechanical protection and must be installed so that they cannot be mechanically damaged from the outside; e.g., installing pumping units with impact protection, attaching shatter protection for glass flasks, etc.
- The devices are designed for an ambient and media temperature of +10 °C to – +40 °C during operation. These ambient and media temperatures must never be exceeded. When conveying/

measuring non-explosive gases, extended gas suction temperatures apply, see chapter: Technical data, media temperature (gas).

2.11 Disposal



NOTICE

Improper disposal of electronic components can result in damage to the environment.

Electronic equipment contains hazardous materials that can damage the environment or human health. Discarded electronic equipment also contains valuable raw materials, which can be recovered if properly disposed of for recycling.

End users are required by law to bring waste electrical and electronic equipment to an approved collection site.


- ⇒ Please properly dispose of electrical waste and electronic components at the end of their service life.
- ⇒ Observe the national regulations on disposal and environmental protection.

3 Product description

Pumping units of the PC 3001 VARIO select series essentially consist of membrane pump, regulated by VARIO drive, a VACUU·SELECT® vacuum controller as well as a cooler with a separator. Coolers are available in different designs. The coolers differ in how they function.

3.1 Basic configuration of pumping unit series

View and basic configuration
PC 3001 VARIO select




Meaning

- | | |
|-----------|--|
| 1 | Power connection with On/Off switch (rocker switch) pumping unit |
| 2 | Separator flask AK, round flask at inlet |
| 3 | Distributor head |
| 4 | Nameplate |
| 5 | Handle |
| 6 | Gas ballast valve |
| 7 | Connections on EK: outlet, coolant |
| 8 | Pressure relief valve |
| 9 | On/Off button VACUU·SELECT® Controller |
| 10 | Chemical membrane pump |
| 11 | VACUU·SELECT® Sensor, installed in pumping unit housing |
| 12 | Emission condenser EK |
| 13 | Vacuum inlet, on rear round flask |
| 14 | Round flask at outlet |

15 VACUU-SELECT® control unit, removable

3.2 Chemistry pumping unit series

Overview of chemistry pumping unit models



Meaning

Chemistry pumping unit	AK	IK	EK	TE	EKP
a PC 3001 VARIO select	•		•		
b PC 3001 VARIO select TE	•			•	
c PC 3001 VARIO select IK		•	•		
d PC 3001 VARIO select EKP	•				•

Product-specific abbreviations

Product-specific abbreviations


AK	Separator flask, installed at inlet or outlet
EK	Emission condenser, installed at outlet
EKP	Peltronic emission condenser®, installed at outlet
IK	Immission condenser, installed at inlet
w/o EK	Without emission condenser
PC	Chemistry pumping unit with type code
TE	Dry ice condenser, dry ice cooler

3.3 Condensers and coolers

3.3.1 Separator/condenser at inlet

Connection on separator flask

Connections on AK




Meaning

1 Vacuum inlet connection IN

Connection and coolant on the inlet condenser

Connections on the
IK



Description

1 Inlet connection vacuum IN

2 Inlet connection coolant IN, e.g., water

3 Coolant outlet connection EX

3.3.2 Condenser at outlet

Connection and coolant on emission condenser

Connections on EK




Meaning

- 1** Outlet connection coolant EX
- 2** Inlet connection coolant IN, e.g. water
- 3** Outlet connection EX

Connection and coolant on the dry ice condenser

Connections on the
TE
PC 3001 VARIO
select TE



Description


- 1** Opening for refrigerant mixture, e.g., dry ice mixture, coolant insert removable for emptying, secured with bayonet fastener

2 Outlet connection EX

3 Vacuum pump connection

Connections on the Peltronic emission condenser

Connections on EKP



1 Outlet connection EX

2 Vacuum pump connection

3 VACUU BUS connection

4 Power connection with on/off switch




⇒ For detailed information and descriptions about the Peltronic emission condenser
-> see the [Peltronic EK instruction manual](#).

3.4 Application example

Evaporation

-> Example
Rotary evaporation




Description

- 1 **PC 3001 VARIO select** vacuum pumping unit
- 2 Exhaust gas hose (diverted into a fume hood)
- 3 Chiller
- 4 Vacuum hose
- 5 Coolant hoses (connected in series)
- 6 Example: Rotary evaporation

Vacuum concentrator

-> Example
Vacuum concentra-
tor



Description

- 1 **PC 3001 VARIO select TE** vacuum pumping unit
- 2 Exhaust gas hose (diverted into a fume hood)
- 3 Vacuum hose
- 4 Vacuum valve: shut-off valve

5 Process example: Vacuum concentrator

4 Installation and connection

4.1 Transport

Products from **VACUUBRAND** are packaged in stable, recyclable packaging.



The original packaging is customized to your product for safe transport.

⇒ If possible, keep the original packaging, e.g. for sending in for repairs.


Incoming goods

- ⇒ Check the delivery immediately after receipt for any transport damage and for completeness.
- ⇒ Promptly report transport damage to the supplier in writing.

Unpacking

-> Example
Pumping unit in original packaging

Glass flask in enclosed box



1. Only lift the device by the designated handles or grip recesses.
2. Remove the connections, such as hose nozzles and screw connections, from the glass flask.
3. Compare the scope of delivery with the delivery note.

4.2 Installation

NOTICE

Condensate can damage the electronics.

A large difference in temperature between the storage location and the installation site can lead to the formation of condensate.

⇒ After receipt or storage, allow your vacuum to acclimatize for at least 3–4 hours before putting it into operation.


Checking the installation conditions

Calibrating the installation conditions

- The device is acclimatized.
- The ambient conditions are met and lie within the limitations of use.
- The pump must be installed on a stable and secure floor, with no other mechanical contact apart from the pump feet.

Installing the vacuum pump

-> Example
Sketch of minimum
distances in laboratory
furniture



- ⇒ Place the vacuum pump on a stable, vibration-free, level surface.
- ⇒ When installing laboratory furniture, maintain the minimum distance of 5 cm (2 in.) to adjacent objects or surfaces.
- ⇒ Avoid heat accumulation and ensure sufficient air circulation, especially in closed housings.

Observing the limitations of use

Ambient conditions


Ambient conditions		(US)
Ambient temperature	10–40 °C	50–104°F
Installation height, max.	2000 m above sea level	6562 ft above sea level
Humidity	30–85 %, non-condensing	
Degree of contamination	2	
Impact energy	5 J	
Protection class (IEC 60529)	IP 20	
Protection class (UL 50E)	Type 1	
Avoid condensate or contamination from dust, liquids and corrosive gases.		

- ⇒ Note the indicated IP protection rating. The IP protection is only guaranteed if the device properly installed and connected.
- ⇒ When connecting the device, always take note of the specifications from the type plate and in the chapter technical data.

4.3 Connection (supply connections)

On the pumping unit, supply connections are provided for vacuum and exhaust, and optionally for gas ballast, ventilation and coolant. Set up the connection for your pumping unit as described in the following examples. Also attach the provided screw fasteners and glass flasks to the condensers.

4.3.1 Vacuum connection (IN)

	CAUTION
	<p>Flexible vacuum hoses can contract during evacuation.</p> <p>Non-fixed, connected components may cause injury or do damage due to the jerky movement (shrinking) of the flexible vacuum hose. The vacuum hose can come loose.</p> <ul style="list-style-type: none"> ➤ Fasten the vacuum hose to the connections. ➤ Fasten connected components. ➤ Measure the flexible vacuum hose so that the maximum shrinkage, i.e. the contraction, is taken into account.


NOTICE

Foreign bodies in the suction line can damage the vacuum pump.

⇒ Prevent particles, liquids or contaminants from being vacuumed or from being able to flow back.

Connecting the vacuum hose

-> Example
Vacuum connection
at inlet IN



1. Join the sealing ring **(a)**, the cap nut **(b)** and hose nozzle **(c)** as pictured.
2. Slide the vacuum hose **(d)** from the apparatus onto the hose nozzle and fasten the vacuum hose, e.g. with a hose clamp.



You will achieve the optimal vacuum for your application if you observe the following points:

- ⇒ Connect the shortest possible vacuum line with the maximum possible cross-section.
- ⇒ Use a vacuum hose with sufficient stability that is designed for the vacuum range used.
- ⇒ Connect hose lines so they are gas tight.

4.3.2 Exhaust connection (OUT)



WARNING


Risk of bursting due to excess pressure in the exhaust line.

Unacceptably high pressure in the exhaust line may cause the vacuum pump to burst or damage seals.

- The exhaust line (outlet, gas outlet) must always be open and free of pressure.
- Always lay the exhaust hose so that it is descending, or take measures to prevent condensate return flow into the vacuum pump.
- Observe the maximum allowable pressures and pressure differentials.

Connecting the exhaust hose

-> Example
Exhaust connection
at outlet EX



1. Join the rubber sealing ring **(a)**, the hose nozzle **(b)** and the cap nut **(c)** as pictured and screw this onto the connection.
2. Slide the exhaust hose **(d)** onto the hose nozzle and, if necessary, route the hose to a fume hood. If necessary, fasten the exhaust hose, e.g. with a hose clamp.

4.3.3 Coolant connection on the condenser


Coolant connection
Inflow and outflow

An emission condenser EK has one connection for coolants. Water or liquid in the circuit of a recirculating chiller, for example, is suitable for cooling.

- In a closed, internal coolant circuit, the pressure should be limited to 3 bar (44 psi).
- A cooling water valve may only be installed in the inflow, the coolant outflow must be open and free of pressure.

Connecting coolant⁵

-> Example
Coolant connection
on EK or IK



1. Fasten both hose nozzles **(a)** with the union nut **(b)** to the condenser as pictured.
2. Attach the hoses for the coolant according to the illustration on the condenser:
IN = Inflow
EX = Outflow
3. Fasten the hoses, e.g. with hose clamps.

⁵ Also applies to immission condenser IK

4.3.4 Dry ice condenser

NOTICE

Damage to the dry ice condenser due to cryogenic substances.

- ⇒ Carry out a visual inspection prior to each use. Glass surfaces must be free from damage, ruptures, cracks and scratches.
- ⇒ Only loosely lay on the cover of the dry ice condenser, thus ensuring pressure equalization between the coolant and the atmosphere.
- ⇒ The coolant can escape from the chiller unexpectedly, e.g., in case of high gas volume.

Cooling with dry ice condenser

Cooling with refrigerant mixtures

The dry ice condenser does not have a coolant connection. For cooling, the dry ice chiller is filled with a refrigerant mixture. These refrigerant mixtures consist of cold to cryogenic media and a liquid for better cooling transfer.

Details refrigerant mixture

-> Example
Refrigerant mixtures

Refrigerant mixtures		
Ethanol/dry ice mixture		
Water/ice mixture		
Brine/ice mixture		
Admissible cooling temperatures		(US)
cold	-18 – -5 °C	-0.4 – 23 °F
very cold	-30 – -18 °C	-22 – -0.4 °F
cryogenic	below -30 °C	less than -22 °F
lowest	-80 °C	-112 °F

Fill the dry ice condenser




CAUTION

Risk of injury when handling cryogenic coolants.

Cryogenic substances can cause frostbite, or 'cold burns', on contact with skin.

- Avoid skin contact and always wear your personal protective equipment, e.g., thermal protective gloves, protective goggles, when handling cryogenic substances.

-> Example
Fill the dry ice con-
denser with refriger-
ant mixture







1. Remove cover **(a)** from dry ice chiller **(b)**.
2. Use your preferred refrigerant mixture **(c)** to fill container **(d)**.
 - Do not overfill the container.
3. Then replace the cover on the dry ice chiller.
 - Only ever lay the cover on loosely, do not secure it.
 - Regularly check the coolant level in the chiller during operation.

Empty the dry ice condenser TE

Before refilling the dry ice condenser with coolant, it may need to be emptied first. Remove the chiller insert (bayonet fastener) and empty it.

-> Example
Chiller insert (bayonet fastener)

	
<p>1. Remove the cover from the chiller.</p>	<p>2. Rotate the chiller insert – bayonet fastener.</p>
	
<p>3. Pull out the chiller insert.</p>	<p>4. Empty the liquid.</p>
<p>5. Refit the empty chiller insert in the dry ice condenser in reverse order.</p>	

4.3.5 Venting connection



DANGER

Explosion risk due to air ventilation.

Depending on the process, a potentially explosive mixture can form during ventilation, or other dangerous situations can occur.

- Never ventilate processes with air in which a potentially explosive mixture can form.
- When flammable substances are present, use only inert gas for ventilation, e.g. nitrogen (max. 1.2 bar/900 Torr abs.).

Vent with ambient air⁶

Position of sensor +
venting valve sec-
tional diagram




For venting with ambient air, nothing needs to be connected to venting valve **(b)** of sensor **(a)**.

⁶ Only applicable to sensors with an integrated venting valve.

Vent with inert gas – connect venting valve⁷

Required connection material: Hose for hose nozzle (\varnothing 4–5 mm), e.g., silicone tube 3/6 mm.

Position of sensor + inert gas venting valve (sectional diagram)



1. Slightly tilt the pumping unit to one side and push hose **(c)** onto the connection of venting valve **(b)**.
2. Route the hose underneath the pumping unit to the outside and connect the inert gas (max. 1.2 bar/900 Torr, abs.).

⁷ Avoid overpressure.

4.3.6 Gas ballast (GB)

Use ambient air as gas ballast




DANGER

Explosion risk due to air as gas ballast.

By using air as a gas ballast, a small amount of oxygen enters the inside of the vacuum pump. Depending on the process, a potentially explosive mixture can form due to oxygen in the air, or other dangerous situations can occur.

- In the presence of flammable substances and for processes in which a potentially explosive mixture can occur, use only inert gas as a gas ballast, e.g. nitrogen (max. 1.2 bar/900 Torr abs.).


Gas ballast valve position



If ambient air is to be used as gas ballast, nothing needs to be connected at the pumping unit; gas ballast valve **(a)**; see also chapter: → **Operation with gas ballast on page 50**

Use of inert gas as gas ballast – OPTION

Prepare the inert gas connection (GB)



- ⇒ Remove the black gas ballast cap and connect a gas ballast adapter in its place.




On request, we can send you connection options and adapters for hose nozzles or small flanges.

4.4 Electrical connection

Connecting the pumping unit to the electrical system

-> Example
Electrical connection
of the pumping unit




1. Plug the socket **(a)** of the power cable into the power connection of the vacuum pump.
2. Connect the power plug **(b)** to the power outlet.
 - ☑ Pumping unit connected to the electrical system.

NOTICE! Lay the power cable so that it cannot be damaged by sharp edges, chemicals or hot surfaces.

Power connections with country code

-> Example
Power plug types



*Diagrams of standard
power connections
with ground contact*

1 UK

2 CN

3 IND

4 US

5 CEE

6 CH

The vacuum pump is delivered ready for use with the matching power plug.

NOTICE!


- ⇒ Use the power plug that matches your power connection.
- ⇒ Do not use multiple-socket power strips as a power connection.
- ⇒ The power plug also works as a circuit breaker. Set up the device so that the plug can be easily disconnected from it.

Connection options for vacuum accessories

The VACUU BUS interface functions as the power supply and control line for vacuum accessories.

1. Connect your accessories to your controller via the VACUU·BUS cable.
2. If necessary, increase the range and the number of connections with a suitable Y-adapter and extension cable.

-> Example
Schematic drawing
of controller with
connected valve and
sensors



Accessories -> see chapter Ordering information

5 Operation

Before putting into operation, make sure that the activities described in the chapter **Installation and power connection** have been carried out properly.


With the exception of the chapter Switching on and off, this instruction manual contains descriptions about the mechanics of a PC 3001 VARIO select series pumping unit.

The operation of an installed vacuum regulator ⁸ and its functions are described in the instruction manual of a **VACUU·SELECT**.

5.1 Switching on

Switching on the pumping unit

Switching on



1. Switch on the rocker switch **(a)** – switch position **I**.
2. Press the ON/OFF button **(b)** on the controller.
 - Display with start screen.
 - After approx. 30 seconds, the process display appears with the control elements in the display of the controller.


⁸ WEB: VACUUBRAND/Produkte/Messgeraete_und_Controller/Vakuum_regeln

5.2 Operation with controller

5.2.1 Operator interface


Operator interface

VACUU·SELECT® with process display



Process display










Pressure display for a process



- 1 Status bar
- 2 Analog pressure display – pressure curve
- 3 Digital pressure display – pressure value (target value, actual value, pressure unit)
- 4 Process display with context functions
- 5 Screen navigation
- 6 Controls for the controller

Controls

Controls
Vacuum controller


Button	Function
 	<p>Start Start an application – only in the process display</p>
 	<p>Stop Stop an application – always possible.</p>
	<p>VENT⁹ – Ventilate system (Option) Button press < 2 Sec. = briefly ventilate, regulation continues.</p>
 	<p>Button press > 2 Sec. = ventilate to atmospheric pressure, vacuum pump is stopped. Button press during ventilation = ventilation is stopped.</p>
 	<p>Favorites Call up Favorites menu.</p>

9 The VENT button is only shown if a ventilation valve is connected or activated.

5.2.2 Operation


Starting the vacuum controller

Start




Stopping the vacuum controller

Stop



Ventilating

Ventilating




5.2.3 Operation with gas ballast

Meaning The supply from the gas ballast (= gas addition) ensures that vapors are not condensed in the vacuum pump; instead, they are emitted out of the pump. This allows greater amounts of condensable vapors to be pumped, which extends the service life. The end vacuum with gas ballast is slightly higher.

Opening/closing the gas ballast valve

Operating the gas ballast valve



- ⇒ Turn the black gas ballast cap in any direction to open or close the gas ballast valve.
- ⇒ If possible, only evacuate the condensable vapors, e.g. water vapor, solvents, etc., when the vacuum pump is at operating temperature and the gas ballast valve is open.
- ⇒ Connect inert gas as the gas ballast to prevent and exclude the formation of potentially explosive mixtures during operation.
- ⇒ Comply with the max. permitted pressure of 1.2 bar/900 Torr abs. at the gas ballast connection.



If the gas volume in the vacuum pump is low, a gas ballast can be eliminated in these cases to increase the solvent recovery rate.


5.3 Switching off (decommissioning)

Switching off the pumping unit

Switching off, e.g. taking the pumping unit out of operation

1. Stop the process and allow the pumping unit to continue running for approx. 30 minutes with open gas ballast or open inlet (IN).
 - Condensate and residual media are rinsed out of the vacuum pump.

NOTICE! Avoid deposits and rinse the condensate out of the pump.



2. Switch off the rocker switch **(a)** – switch position 0.
 - Pumping unit switched off.
3. Disconnect the power plug.
4. Disconnect the pumping unit from the equipment.
5. Empty the glass flasks.
6. Check the pumping unit for any damage and soiling.

5.4 Storage

Storing the pumping unit

1. Clean the The pumping unit if it is dirty.
2. Recommendation: Carry out preventative maintenance before you put The pumping unit into storage. Especially if it has been in operation for more than 15,000 hours.
3. Seal off the suction and exhaust line, e.g. with the transport caps.
4. Pack up The pumping unit so that it is protected from dust; possibly include a drying agent.
5. Store The pumping unit in a cool and dry location.

NOTICE! If for operational reasons damaged parts are stored, these should be clearly marked as non-operational.

6 Troubleshooting

6.1 Technical assistance

To find and eliminate errors, use the table → **Error – cause – corrective measure on page 53.**

For technical assistance or in case of faults, please contact our **Service** department.



The device should only be operated in perfect technical condition.

- ⇒ Adhere to the recommended maintenance intervals to ensure a properly functioning system.
- ⇒ Send defective devices to our Service department or your specialist retailer for repairs.

6.2 Error – cause – corrective measure

Error	Cause	Corrective measure	Personnel
Measurements deviate from the reference standard	Sensor dirty. Humidity in sensor. Sensor defective. Sensor not measuring correctly.	Clean sensor measuring chamber. Allow sensor measuring chamber to dry, e.g. by draining. Calibrate sensor with reference measuring device. Replace defective components.	Qualified employee
Sensor does not transmit measurement	No voltage applied. VACUU BUS plug connection or cable defective or not inserted.	Check the VACUU BUS plug connection and cable to the controller.	Operator
Sensor does not transmit measurement	Sensor defective.	Replace defective components.	Qualified employee
Ventilation valve does not switch	No voltage applied. VACUU BUS plug connection or cable defective or not inserted. Ventilation valve dirty.	Check the VACUU BUS plug connection and cable to the controller. Clean the ventilation valve.	Operator

Error	Cause	Corrective measure	Personnel
Ventilation valve does not switch	Ventilation valve sensor defective.	Use another external ventilation valve if necessary. Replace defective components.	Qualified employee
Vacuum pump does not start	Pumping unit switched off. Power plug not connected properly or pulled out. VACUU BUS plug connection or cable defective or not inserted. Excess pressure in the exhaust line.	Switch on Pumping unit. Check the power connection and cable. Check the VACUU BUS plug connection and cable to the controller. Open the exhaust line. Ensure that nothing is blocking the line.	Operator
Vacuum pump stopped Vacuum pump does not start	Motor overloaded. Motor overheating. Thermal protection activated.	Check coolant connection. Ensure coolant supply. Allow the motor to cool down. Manually reset fault: -> Disconnect the pumping unit from the power -> Eliminate the error cause -> Restart the pumping unit	Qualified employee
No suction capacity or very little	Leak in the suction line or on the apparatus. Condensate collection bottle not installed correctly. Condensate in the vacuum pump. Open gas ballast Gas ballast cap porous or no longer present.	Check suction line and apparatus for any leaks. Check condensate collection bottle and install correctly. Check apparatus for leaks. Let the vacuum pump run for a few minutes with open suction intake port. Close gas ballast	Operator

Error	Cause	Corrective measure	Personnel
		Check the gas ballast cap. Replace defective components.	
No suction capacity or very little	Deposits in the vacuum pump. Membrane or valves defective. High amount of vapor build up in process.	Clean and test pump heads. Replace the membrane and valves. Check process parameters.	Qualified employee
No suction capacity or very little	Vacuum line too long.	Use vacuum lines with a larger cross-section.	Resp. specialist
Display off	Pumping unit switched off. Power plug not connected properly or pulled out. VACUU BUS plug connection or cable defective or not inserted. Controller switched off or defective.	Switch on Pumping unit. Check the power connection and cable. Check the VACUU BUS plug connection and cable to the controller. Replace defective components.	Operator
Condenser (cooler) defective	Mechanically damaged.	Send in.	Resp. specialist
Loud operating noises	No hose installed.	Check hose and install correctly.	Operator
Loud operating noises	Exhaust line open. Glass flask on EK missing. Membrane torn or membrane spring washer loose. Ball bearings defective.	Check exhaust line connections. Connect exhaust line to a suction or extraction system. Install glass flask Service vacuum pump and replace defective parts, or send device in for repair.	Qualified employee

7 Cleaning and maintenance



WARNING

Danger due to electrical voltage.

- Switch off the device before cleaning or maintenance.
- Disconnect the power plug from the power outlet.



Danger due to contaminated components.

When pumping dangerous media, hazardous materials can adhere to interior pump components.

If this case applies to you:

- ⇒ Wear your personal protective equipment, e.g. safety gloves, eye protection and, if required, a respirator.
- ⇒ Decontaminate the vacuum pump before you open it. If required, have the vacuum pump decontaminated by an external service provider.
- ⇒ Take safety precautions when handling hazardous materials, in accordance with your operating instructions.

NOTICE

Carrying out work in an improper manner may cause damage.

- ⇒ Have Maintenance tasks carried out by a qualified specialist or at least a trained person.
- ⇒ Before carrying out the first Maintenance, please read through all action instructions to obtain an overview of the service activities required.

7.1 Information on service activities

Recommended maintenance interval ¹⁰

Maintenance intervals

Maintenance intervals	As required	15,000 h
Replace membranes		x
Replace valves		x
Clean or replace PTFE molded hose	x	
Replace pressure release valve on EK	x	
Clean pumping unit	x	

Recommended auxiliary equipment

->Example Recommended auxiliary equipment for cleaning and maintenance



Meaning

No.	Auxiliary equipment
1	Mat for round flasks
2	Glass pipette
3	Protective gloves
4	Chemical resistant container and funnel

¹⁰ Recommended maintenance interval according to hours of operation and under normal operating conditions; depending on the setting and field of application, we recommend carrying out cleaning and maintenance as needed.

Tools needed for maintenance

-> Example Tools




Meaning

No.	Tool	Size
1	Flat-head screwdriver Opening hose clamps	Size 1
2	Torx screwdriver Screw fasteners for EK or IK counterholder	TX10
3	Open-end wrench Union nut M14 Rotate elbow fitting	SW17 SW14
4	Phillips-head screwdriver Screw fasteners for TE or EKP bracket	Size 2
5	Flat-nose pliers Closing hose clamps	
6	Hexagon socket wrench Screw fasteners for handle Screw fasteners for housing cover Fixing screws for sensor	Size 5 Size 4 Size 3
7	Seal set PC 3001 #20696828 Membrane wrench Membrane Valves	SW46

7.2 Cleaning

This chapter does not describe how to decontaminate the product. Simple cleaning and care measures are described here.

⇒ Before cleaning, switch off the pumping unit.

	CAUTION
Risk of burning due to hot surfaces	
An elevated exhaust gas temperature can lead to hot surfaces on the instrument and on attached components, such as glass flasks. The temperatures that are produced during operation may cause burns.	
<ul style="list-style-type: none">➤ Use protection against accidental contact, especially when the exhaust temperature is persistently high.➤ Allow the instrument to cool before you empty the glass flask or begin performing maintenance tasks.➤ For tasks that must be performed during operation, use your personal protection equipment, e.g., heat-resistant safety gloves.	

7.2.1 Housing surface

Cleaning the surface




Clean soiled surfaces with a clean, slightly damp cloth. We recommend water or a mild soap solution to moisten the cloth.

7.2.2 Emptying the glass flask

Removing and emptying the glass flask

-> Example
Emptying the glass
flask



1. Open the joint clamp and remove the glass flask.
2. Empty the glass flask into a suitable container, e.g. a chemical-resistant canister.
3. Reattach the glass flask (separator) to the condenser with the joint clamp.




Depending on the application, the collected liquid can either be recycled or properly disposed of.

7.2.3 Clean the sensor and venting valve


In case of incorrect measurements or malfunctions that indicate contamination of the sensor and/or venting valve, we recommend cleaning the sensor and the venting valve. Cleaning is also recommended prior to re-calibration.

Remove the sensor

-> Example
Remove the sensor




1. Switch the pumping unit off and unplug the power plug.
2. Remove the vacuum controller from the housing and disconnect any connected VACUU-BUS connectors.
3. Remove the glass flasks and place them on suitable stands.



4. Close the hose nozzle of the chiller and carefully turn the pumping unit upside down.
5. Open the union nut on the sensor, open-end wrench SW17, and pull off the molded hose.

-> Example
Remove the sensor




6. Undo the screws, hex key size 3



7. Disconnect the VACUUBUS connector at the bottom and remove the sensor.

Clean the sensor

-> Example
Clean the measurement chamber and the venting valve



1. Use a pipette to fill a small amount of solvent, e.g., benzene, into the openings.



2. Leave the solvent to act for a few minutes before pouring it off.

3. Repeat until the solvent is free from contaminants.


4. Let the interior of the sensor air dry or dry under vacuum.

Install the sensor


-> Example
Fit the sensor



1. Connect the VACUU-BUS connector and place the sensor on the holder.



2. Insert the screws and screw them in until they are finger-tight, hex key size 3.



3. Push the molded hose onto the connection and screw on the union nut until it is finger-tight; open-end wrench SW17.



4. Turn the pumping unit right side up and connect the cables: VACUU-BUS, power plug.

5. Secure the glass flasks with the joint clamp.

6. Switch on the pumping unit and the vacuum controller.

NOTICE! Re-calibrate the sensor if incorrect values are displayed -> see vacuum controller manual.

7.2.4 Cleaning or replacing PTFE molded hoses

Maintenance provides an opportunity to check the components of the pumping unit, including the hosing.


- ⇒ Clean the inside of highly soiled molded hoses, e.g. with a pipe cleaner or similar.
- ⇒ Replace cracked or defective molded hoses.

7.3 Vacuum pump maintenance

7.3.1 Maintenance items

Positions to be maintained

-> Example
Pump head maintenance



Description

Maintenance items

- 1** Housing cover, power supply side
- 2** Housing cover with gas ballast

- ⇒ Perform maintenance on the pump heads one after the other.
- ⇒ Always replace all diaphragms and valves of the pump heads as described in the graphic description for pump head **(1)**.




Straightforward maintenance due to split work steps.

- ⇒ On one pump head pair, first replace the diaphragms.
- ⇒ Then change the inlet/outlet valves.
- ⇒ Then repeat these steps for the next pump head.

Exploded drawing of pump head (example)

-> Example
Exploded drawing of
pump head



Description

Valve maintenance

- 1 Screw fittings
- 2 Housing cover
- 3 Valves


Diaphragm maintenance

- 4 Head cover
- 5 Diaphragm clamping disc with square head screw
- 6 Diaphragms
- 7 Diaphragm support disc
- 8 Spacer discs, max. 4
- 9 Pumping unit


7.3.2 Change the diaphragms and valves

Preparation

-> Example
Prepare for maintenance



1. Switch the pumping unit off and unplug the power plug.



2. Remove the glass flasks and connected hoses (coolant, vacuum).

-> Example
Remove the EK (IK)



3. Unscrew the screws from the counterhold;
Torx screwdriver TX10.



4. Remove the counterhold and set it aside with the screws.

-> Example
Remove the EK (IK)



5. Unscrew the union nut and pull off the molded hose.


6. Remove the chiller.



7. Set the chiller down securely so that no liquid can escape.

Remove the TE or EKP

-> Example
Remove the dry ice condenser (TE) or Peltronic vapor condenser (EKP)



⇒ Unscrew the 2 screws; Phillips screwdriver size 2.



The chillers TE and EKP are secured with retaining plates.

⇒ On these chillers, only loosen the screws in the retaining plates on the pumping unit.

Disassemble the device and housing sections

-> Example
Remove the housing
sections on the left



1. Loosen the union nut; open-end wrench SW17.



2. Rotate the angled screw fitting to one side by a quarter turn; open-end wrench SW14.



3. Loosen the screw fitting on the handle; hex key size 5.



4. Carefully lay the pumping unit on its side.



5. Open the hose clip; flat-head screwdriver size 1.




6. Unscrew the screw fittings; hex key size 4.


-> Example
Remove the housing
sections on the left



7. Lift off the housing cover and pull off the molded hose.



8. Check for adhered valves and place the housing cover with the screw fittings to one side.



9. Make a note of the position of the head covers.




10. Remove the head covers.

NOTICE! Valves must be correctly positioned, otherwise the vacuum pump will not generate a vacuum.

Replace diaphragms

-> Example
Diaphragm replacement



1. Fold the diaphragm up at the sides.





2. Carefully position the diaphragm wrench on the diaphragm support disc.



3. Use the secured diaphragm wrench to unscrew the assembly.




4. Lift the diaphragm, along with all the parts, out of the vacuum pump.



- Never drop spacer discs into the aluminum housing.
- Check for any spacer discs adhering to the connecting rod.
- Keep the spacer discs. It is essential to reinsert the same number of spacer discs.

5. Disassemble the assembly and take a new diaphragm; service kit MD 1C.



- Make sure that you install the membrane correctly, with the coated, light side on top.
- Ensure that it is correctly positioned on the square bolt.




6. Assemble the diaphragm assembly and ensure its correct positioning on the square head.




7. Use the correct number of spacer discs.




8. Secure the diaphragm assembly inside the diaphragm wrench.



9. Hold the spacer discs firmly and place the assembly on the connecting rod thread.



10. Tighten the assembly until finger-tight, using the diaphragm wrench.




11. Repeat for the second diaphragm.

Change valves


-> Example
Valve replacement




1. Carefully clean dirty head covers and




2. housing covers with a cloth.



3. Place both head covers in the correct position.



4. Position the new valves and align them; service kit MD 1C.



Cutout view from above: correct valve positioning.



5. Once all valves are positioned correctly, initially insert the molded hose.

-> Example
Valve replacement



6. Position the housing cover so that it is level and screw in the screw fittings; hex key size 4, tightening torque 6 Nm.

Assemble the device and housing sections

Before restarting the pumping unit, all previously removed parts of the device and housing must be fixed back in place.

-> Example
Mount the device and
housing sections




1. Set the pumping unit up-right.



2. Rotate the angled screw fitting back into the molded hose by a quarter turn.



3. Screw on the union nut until it is finger-tight; open-end wrench SW17.



4. Secure the handle; hex key size 5.



5. Secure open hose clips using flat nose pliers.



6. Push on the molded hose and screw on the union nut until it is finger-tight.



7. Secure the counter holder; Torx screwdriver TX10
For EKP or TE, please screw on the retaining plate; Phillips screwdriver size 2.



8. Secure the glass flasks with the joint clamp.

Diaphragm and valve replacement, next pump head

-> Example
Second pump head
maintenance



1. Turn the The pumping unit to the other side.
2. Repeat the above steps for diaphragm and valve replacement.

If the maintenance work is fully completed:

- ⇒ Connect the hosing for operation.
- ⇒ Connect the pumping unit to the power supply.
 - ☑ Pumping unit ready for restart.
 - ☑ Without reconnection -> Pumping unit prepared for storage.

8 Annex

8.1 Technical data

Product description


Product names

Chemistry pumping unit series

PC 3001 VARIO select	PC 3001 VARIO select IK
PC 3001 VARIO select TE	PC 3001 VARIO select EKP

Technical data

Technical data

Ambient conditions		(US)
Ambient temperature	10–40 °C	50–104°F
Installation height, max.	2000 m above sea level	6562 ft above sea level
Humidity	30–85 %, non-condensing	
Degree of contamination	2	
Impact energy	5 J	
Protection class (IEC 60529)	IP 20	
Protection class (UL 50E)	Type 1	
Avoid condensate or contamination from dust, liquids and corrosive gases.		
Operating conditions		(US)
Operating temperature	10–40 °C	50–104 °F
Storage/transport temperature	-10–60 °C	14–140 °F
Maximum permissible media temperature (gas) of non-explosive atmospheres:		
short-term	80 °C	176 °F
Continuous operation	40 °C	113 °F
ATEX conformity	II 3/- G Ex h IIC T3 Gc X Internal Atm. Only Tech. File: VAC-EX02	
Maximum permissible media temperature (gas) of  atmospheres:		
short-term	40 °C	104 °F
Continuous operation	40 °C	104 °F
Connections		
Vacuum, inlet IN	Hose nozzle DN 6/10	
Gas ballast GB	Gas ballast valve, manual	

Inert gas adapter – OPTION	Small flange GB NT KF DN 16 Hose nozzle GB NT DN 6/10	
Ventilation valve (ventilation with inert gas) – OPTION	Silicon rubber hose 3/6	
Coolant EK (+IK)	2x (+2x) hose nozzle DN 6/8	
Exhaust, outlet EX	Hose nozzle DN 8/10	
Cold-device plug	+ power connection CEE, CH, CN, UK, IN, US	
Plug connector	VACUU·BUS [®]	
Electrical data		
Rated voltage	100–230 VAC ±10 %	
Power frequency	50/60 Hz	
Overvoltage category	II	
Rated current at 50 Hz	1.6–0.7 A	
Power, max.	0.16 kW	
Interface	VACUU·BUS [®]	
Power cable	2 m	
Device fuse on motor circuit board	1x 1.1 AT (VACUU·BUS) 1x 7 AF	
Vacuum data		
		(US)
Inlet pressure / outlet pressure / differential pressure, absolute	1.1 bar	825 Torr
Pressure at gas connections, absolute max.	1.2 bar	900 Torr
Sensor	integrated	integrated
Measurement principle	Ceramic membrane (aluminum oxide), capacitive, gas-type independent, absolute pressure	
Measurement accuracy	±1 mbar/hPa/Torr, ±1 digit (after adjustment, constant temperature)	
Upper measurement limit	1080 mbar	810 Torr
Lower measurement limit	0.1 mbar	0.1 Torr
Temperature drift	< 0.15 mbar/K	0.11 Torr/K
Max. pumping speed	2.0 m ³ /h	1.18 cfm
End vacuum, absolute	2.0 mbar	1.5 Torr
End vacuum with GB, absolute	4 mbar	3 Torr
Number of cylinders/steps	4/3	

Weights* and dimensions (L x W x H)		(US)
PC 3001 VARIO select	303 mm x 306 mm x 400 mm	12.05 in x 11.93 in x 15.75 in
Weight*	8.2 kg	18.08 lb
PC 3001 VARIO select TE	300 mm x 341 mm x 493 mm	11.81 in x 13.43 in x 19.41 in
Weight*	8.7 kg	19.18 lb
PC 3001 VARIO select IK	309 mm x 312 mm x 400 mm	12.17 in x 12.28 in x 15.75 in
Weight*	8.8 kg	19.4 lb
PC 3001 VARIO select EKP	300 mm x 370 mm x 400 mm	11.81 in x 14.57 in x 15.75 in
Weight*	11.8 kg	26.01 lb
* without cable		

Other specifications

Sensor type	VACUU·SELECT Sensor
Controller	VACUU·SELECT
Volume of condensate collection tanks	500 ml each
Sound pressure level at 1,500 min ⁻¹ /62% (VARIO)	42 dBA ±3

8.2 Wetted materials

Wetted materials

Component	Wetted materials
Pump	
Housing cover	PTFE
Head cover	ETFE carbon fiber reinforced
Diaphragm clamping disc	ETFE carbon fiber reinforced
Diaphragms	PTFE
Valves	FFKM
Pumping unit	
Inlet	PPS (IK: PP)
Outlet	PET (PC 3001 without EK: PTFE carbon fiber reinforced)
Hoses	PTFE
Hose fitting	ETFE, ECTFE
O-ring on separator	Fluoroelastomer

Pressure relief valve at the vapor condenser	Silicone rubber, PTFE film
Distributor head (inlet)	PPS glass fiber reinforced, PP (blind plate)
Condenser IK, EK, TE	Borosilicate glass
Round bottom flask	Borosilicate glass
Peltronic vapor condenser	ETFE, ECTFE, PP, PA
Silencer	PBT, PVF, rubber
VACUU-SELECT Sensor	
Vacuum sensor	Aluminum oxide ceramic, gold-coated
Measuring chamber	PPS
Small flange OPTION	PP
Seal on sensor	Chemical-resistant fluoroelastomer
Hose nozzle	PP
Seal on ventilation valve	FFKM

8.3 Rating plate

Specifications from the type plate

i

⇒ In case of error, take note of the type and serial number from the type plate.

⇒ When contacting our service department, provide the type and serial number from the type plate. This way we can provide you with targeted support and consultation for your product.

Pumping unit type plate, general

-> Example Type plate details

The diagram shows a rectangular type plate for a VACUUBRAND VACUUM PUMPING UNIT. The plate contains the following information and symbols:

- Model/year of manufacture/month:** Indicated by a callout line pointing to the top left area.
- Manufacturer:** VACUUBRAND GMBH + CO KG, 97...77 Wertheim · Alfred - Zippe - Str. 4, Made in Germany. Indicated by a callout line pointing to the top right area.
- Serial number:** Ser.No. / ...
- Suction capacity:** max. m³/h
- Ultimate vacuum:** ... mbar
- Protection class:** IP
- Supply voltage:** ... V, ... Hz, max. ... A
- ATEX specification*:** Indicated by a callout line pointing to the bottom left area.
- Option: Approval/markings/symbols:** Indicated by a callout line pointing to the bottom right area, which includes a CE mark, a 40°C temperature limit symbol, and a warning symbol.

8.4 Order data

Ordering information
for pumping units

Chemistry pumping unit series	*Order no.
PC 3001 VARIO select	2070020x
PC 3001 VARIO select TE	2070022x
PC 3001 VARIO select IK	2070026x
PC 3001 VARIO select EKP	2070024x

* Order no. depends on power cord CEE, CH, UK, US, CN, IN

Order data acces-
sories

Accessories	Order No.
Vacuum hose DN 6 mm (l = 1000 mm)	20686000
Vacuum hose DN 8 mm (l = 1000 mm)	20686001
Coolant valve VKW-B	20674220
Ventilation valve VBM-B	20674217
Level sensor	20699908
VACUU·SELECT Sensor	20612881
VSK 3000	20640530
DAkKS calibration with first delivery	20900214
DAkKS recalibration	20900215

Order data replace-
ment parts

Replacement parts	Order No.
Hose nozzle 6 curved	20639948
Hose nozzle DN 6/10	20636635
Small flange KF DN 16	20635008
Extension cable VACUU·BUS, 0.5 m	20612875
Extension cable VACUU·BUS, 2 m	20612552
Extension cable VACUU·BUS, 10 m	22618493
Spherical joint clamp VA KS35/25	20637627
Glass flask/round flask 500 ml	20638497
PA knurled nut M14x1 (union nut)	20637657
PA clamping ring D10 (seal)	20637658
Emissions condenser EK, complete	On request
Dry ice condenser TE	On request
Immissions condenser IK	On request
Peltronic emissions condenser EKP	20636298
Anti-rotation protection D17x17.5	20635113
Gas ballast cap	20639223

Power cable	CEE	20612058
	CH	20676021
	CN	20635997
	IN	20635365
	UK	20612065
	CEE	20612058



⇒ VACUUBRAND > Support > Repair instructions > [Chemistry pumping units](#).

Supply sources

International agents and dealers

Purchase original accessories and original replacement parts from a branch office of VACUUBRAND GMBH + CO KG or from your local dealer.



- ⇒ Information about the complete range of products can be found in the current product catalog.
- ⇒ Your local dealer or VACUUBRAND [sales office](#) is available for orders, questions about vacuum regulation and the ideal accessories.

8.5 Service information

Use the comprehensive range of services of **VACUUBRAND GMBH + CO KG**.

Detailed overview of services

Range of services

- Product consultation and practical solutions,
- Quick delivery of replacement parts and accessories,
- Professional maintenance,
- Prompt repair process,
- On-site service (on request),
- Calibration (DAkkS accreditation),
- With clearance certificate: Returns, disposal.

Additional information can be found on our website: www.vacuubrand.com.

Service procedure



Follow the description at: VACUUBRAND > Support > [Service](#)



Reduce downtimes, speed up processing. When contacting our service department, please have the required information and documents ready.

- ⇒ Your order can be assigned quickly and easily.
- ⇒ Risks can be excluded.
- ⇒ A short description, photos or diagnostics data help to narrow down the error.

8.6 EU declaration of conformity

EU-Konformitätserklärung EC Declaration of Conformity Déclaration CE de conformité



Hersteller / Manufacturer / Fabricant:

VACUUBRAND GMBH + CO KG · Alfred-Zippe-Str. 4 · 97877 Wertheim · Germany

Hiermit erklärt der Hersteller, dass das Produkt konform ist mit den Bestimmungen der Richtlinien:

Hereby the manufacturer declares that the device is in conformity with the directives:

Par la présente, le fabricant déclare, que le dispositif est conforme aux directives:

2006/42/EG (M-RL), 2014/34/EU (ATEX-RL), 2014/30/EU (EMV-RL),
2011/65/EU, 2015/863 (RoHS-2)

Pumpstand / Pumping unit / Groupe de pompage:

Typ / Type / Type: **PC 3001 VARIO select / PC 3001 TE VARIO select /
PC 3001 EKP VARIO select / PC 3001 IK VARIO select**

Artikelnummer / Order number / Numéro d'article: **20700200, 20700201,
20700202, 20700203, 20700207 / 20700220, 20700223 / 20700245 / 20700265**

Seriennummer / Serial number / Numéro de série: Siehe Typenschild / See rating
plate / Voir plaque signalétique

Angewandte harmonisierte Normen / Harmonized standards applied / Normes
harmonisées utilisées: DIN EN ISO 12100:2011, DIN EN 1012-2:2011,
DIN EN 61010-1:2020, IEC 61010-1:2010 (Ed. 3), DIN EN 1127-1:2019,
DIN EN ISO 80079-36:2016, DIN EN 61326-1:2013, DIN EN IEC 63000:2019

Bevollmächtigter für die Zusammenstellung der technischen Unterlagen / Person
authorised to compile the technical file / Personne autorisée à constituer le
dossier technique:

Dr. Constantin Schöler · VACUUBRAND GMBH + CO KG · Germany

Ort, Datum / place, date / lieu, date: Wertheim, 08.06.2022

(Dr. Constantin Schöler)

*Geschäftsführer / Managing Director /
Gérant*

ppa.

(Jens Kaibel)

*Technischer Leiter / Technical Director /
Directeur technique*

VACUUBRAND GMBH + CO KG

Alfred-Zippe-Str. 4
97877 Wertheim

Tel.: +49 9342 808-0

Fax: +49 9342 808-5555

E-Mail: info@vacuubrand.com

Web: www.vacuubrand.com



www.vacuubrand.com/certificates

8.7 UKCA conformity declaration

UK Declaration of Conformity



Manufacturer:

VACUUBRAND GMBH + CO KG · Alfred-Zippe-Str. 4 · 97877
Wertheim · Germany

Hereby the manufacturer declares that the device is in conformity with the directives:

- Supply of Machinery (Safety) Regulations 2008 (S.I. 2008 No. 1597, as amended by S.I. 2019 No. 696)
- Electromagnetic Compatibility Regulations 2016 (S.I. 2016 No. 1091, as amended by S.I. 2019 No. 696)
- The Equipment and Protective Systems Intended for Use in Potentially Explosive Atmospheres Regulations 2016 (S.I. 2016 No. 1107, as amended by S.I. 2019 No. 696)
- The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (S.I. 2012 No. 3032)

Pumping unit

Type: **PC 3001 VARIO select / PC 3001 TE VARIO select /
PC 3001 EKP VARIO select / PC 3001 IK VARIO select**

Order number: **20700200, 20700201, 20700202, 20700203, 20700207 /
20700220, 20700223 / 20700245 / 20700265**

Serial number: See rating plate

Designated standards applied: EN ISO 12100:2010, EN 1012-2:1996+A1:2009,
EN 61010-1:2010+A1:2019, EN 1012-2:1996+A1:2009, EN 61326-1:2013,
EN 1127-1:2019, EN ISO 80079-36:2016, EN IEC 63000:2018

Person authorised to compile the technical file: Dr. Constantin Schöler ·
VACUUBRAND GMBH + CO KG · Germany

Place, date: Wertheim, 08.06.2022

(Dr. Constantin Schöler)

*Geschäftsführer / Managing Director /
Gérant*

ppa.

(Jens Kaibel)

*Technischer Leiter / Technical Director /
Directeur technique*

VACUUBRAND GMBH + CO KG

Alfred-Zippe-Str. 4
97877 Wertheim

Tel.: +49 9342 808-0

Fax: +49 9342 808-5555

E-Mail: info@vacuubrand.com

Web: www.vacuubrand.com

8.8 Certificate (CUS)

Certificate



Certificate no.

CU 72200438 01

License Holder:
 VACUUBRAND GMBH + Co. KG
 Alfred-Zippe-Str. 4
 97877 Wertheim
 Germany

Manufacturing Plant:
 VACUUBRAND GMBH + Co. KG
 Alfred-Zippe-Str. 4
 97877 Wertheim
 Germany

Test report no.: USA- 32080512 001 **Client Reference:** Dr. Wollschläger
Tested to: UL 61010-1:2012 R7.19
 CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1

Certified Product: Diaphragm vacuum pump system for laboratory use **License Fee - Units**

Model Designation : PC 3001 y zzzzz; VP 2 autovac 7
 y = basic, VARIOpro, VARIOpro TE, VARIOpro IK, VARIO select, VARIO select TE, or VARIO select IK
 z = blank, A-Z, +, -

Input voltage: 100 - 230 V AC; 50/60Hz
Input current: 1.6 - 0.7 A (max)
Protection Class: I

Appendix: 1, 1 - 7

Licensed Test mark:



Date of Issue (day/mo/yr)
 28/02/2020

TUV Rheinland of North America, Inc., 12 Commerce Road, Newtown, CT 06470, Tel (203) 426-0888 Fax (203) 426-4009

Keyword Index

A

Abbreviations used	10
Action instructions (image description)	10
Additional symbols	9
Ambient conditions	33
Application example rotary evaporation	29
ATEX device category	22
ATEX device category and peripheral devices	22

C

Cleaning, general	59
Connecting the exhaust hose	35
Connecting the outlet	35
Connections on EKP	28
Controls for vacuum controller	48
Coolant connection	36
Copyright	6

D

Dangers associated with cryogenic materials	21
Diaphragm change	70
Disassemble the device and housing sections	68
Disposal	23
Dry ice condenser	25
Dry ice cooler	25

E

Emission condenser	25
Empty the dry ice condenser TE	38
Error – cause – corrective measure .	55
Example of use: vacuum concentrator	29
Explanation of safety symbols	9
Explanation of usage/operating conditions X	22

F

Fill the dry ice condenser	38
Fine vacuum	12

I

Immission condenser	25
---------------------------	----

Improper use	14
Inlet condenser connections	26
Installing the vacuum pump	32
Instruction modules	7

L

Labels and signs	21
Looking out for dangers during ventilation	20

M

Maintenance interval	57
Measuring chamber	81
Misuse	14
Modular instruction manuals	7

O

Observe stability	19
Operator interface	47
Operator responsibilities	15
Operator safety measures	17
Overheating protection, blockage protection	21
Overview of chemistry pumping unit models	25

P

PC 3001 VARIO select	25
PC 3001 VARIO select EKP	25
PC 3001 VARIO select IK	25
PC 3001 VARIO select TE	25
Peltronic emission condenser	25
Peltronic® emission condenser	28
Personnel responsibilities	15
Possible residual energies	20
Prepare for maintenance	66
Presentation conventions	8
Presentation of action instructions .	10
Pressure display	47
Preventing backlog in the exhaust line	19
Preventing overheating	20
Preventing sources of ignition	22
Process display	47
Product description	78
Product-specific abbreviations	25
Product-specific terms	12

Proper use.....	13
Pump head exploded drawing	65
Pump head maintenance.....	64

Q

Qualification description	15
Quality standards and safety.....	16

R

Recommended auxiliary equipment for cleaning and maintenance	57
Remove EK (IK) (example).....	66
Remove EKP (Peltronic vapor con- denser)	67
Remove TE (dry ice condenser)	67
Restart procedure	21
Rough vacuum.....	12

S

Safety instructions	13
Separator flask	25
Separator flask connections.....	26
Surface temperatures	20
Switching on	46
Switching on the pumping unit.....	46

T


Target groups	15
Technical data	78

V

Vacuum connection at inlet.....	34
Valve replacement.....	73, 74
Valves in the pump head.....	65
Vent inert gas.....	41
Vent with ambient air.....	40

W

Warning messages	8
Wetted materials	80
Who-does-what matrix	16



[VACUUBRAND > Support > Manuals](#)

Manufacturer:

VACUUBRAND GMBH + CO KG

Alfred-Zippe-Str. 4

97877 Wertheim

GERMANY

Head office: +49 9342 808-0

Sales: +49 9342 808-5550

Service: +49 9342 808-5660

Fax: +49 9342 808-5555

E-mail: info@vacuubrand.com

Web: www.vacuubrand.com